



*Republic of the Philippines*  
Cordillera Administrative Region  
Provincial Local Government Unit of Ifugao  
**BIDS AND AWARDS COMMITTEE**  
3<sup>rd</sup> Floor PEO Bldg., Capitol Compound  
Lagawe, Ifugao  
Email Ad.: bacgoodsifugao13@gmail.com



## INVITATION TO BID

### Publication No. 60-2022-G (First Publication)


#### For Medical Equipment

1. The **Provincial Government of Ifugao**, through the **General Fund** intends to apply the sum of **One Million Two Hundred Eighty-Five Thousand Pesos only (PhP 1,285,000.00)** being the ABC to payments under the contract for **Publication No. 60-2022-G**. Bids received in excess of the ABC shall be automatically rejected at bid opening. **(Please see attached “ANNEX 1” for the item/s to be bided).**
2. The **Provincial Government of Ifugao** now invites bids for the above Procurement Project. Delivery of the Goods is required **within thirty (30) calendar days from receipt by the winning bidder of the Purchase Order**. Bidders should have completed, within **the last five (5) years** from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary “pass/fail” criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
  - a. Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.
4. Prospective Bidders may obtain further information from the Bids and Awards Committee (BAC) of the Provincial Government of Ifugao and inspect the Bidding Documents at the address given below during **8:00 o'clock AM to 5:00 o'clock PM**.
5. A complete set of **Bidding Documents** may be purchased by interested Bidders at the Bids and Awards Committee Office, 3<sup>rd</sup> Floor, PEO Bldg., Capitol, Lagawe, Ifugao on **November 29, 2022 to December 20, 2022** upon presenting a proof of payment in person, of a non-refundable bidding fee of **Five Thousand Pesos (PhP 5,000.00)**.
6. The Bids and Awards Committee (BAC) of the Provincial Government of Ifugao will hold a Pre-Bid Conference on **December 06, 2022 at 10:00 A.M.** at the Bids and Awards Committee Office, 3<sup>rd</sup> Floor, PEO Bldg., Capitol, Lagawe, Ifugao, which shall be open to prospective bidders.
7. Bids must be duly received by the BAC Secretariat through manual submission at the Bids and Awards Committee Office, 3<sup>rd</sup> Floor, PEO Bldg., Capitol, Lagawe, Ifugao on or before **09:30 A.M. basing on the BAC OFFICE - Wall Clock of December 20, 2022**. Late bids shall not be accepted.
8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB** Clause 14.

9. Bid Opening shall be on **December 20, 2022 at 10:00** in the morning at the address stated above. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
10. The **Provincial Government of Ifugao** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
11. For further information, please refer to:

*The Secretariat on Goods  
Bids and Awards Committee  
bacgoodsifugao13@gmail.com  
0965-407-3198*

*Date of Posting: November 29, 2022*

  
**CARMELITA B. BUYUCCAN**  
BAC Chairperson

| "ANNEX 1" |   |     |               |           |            |
|-----------|---|-----|---------------|-----------|------------|
| ITEM No.  |   | QTY | Unit of issue | Unit Cost | Total Cost |
|           | <b>MEDICAL EQUIPMENT</b>  |     |               |           |            |
| 1         | <b>CROSSMATCHING ANALYZER</b>   | 1   | UNIT          |           |            |
|           | with ID Dispensers  |     |               |           |            |
|           | ID system with a 12 ID card capacity  |     |               |           |            |
|           | Automatic balance control at the centrifuge head  |     |               |           |            |
|           | Monitored by microprocessor   |     |               |           |            |
|           | very quiet  |     |               |           |            |
|           | RPM time are functional and displayed (LCD) in English  |     |               |           |            |
|           | speed: 1030-1050 RPM  |     |               |           |            |
|           | Weight: 6.2-6.5kg (centrifuge head included)  |     |               |           |            |
|           | <b>ID Incubator 37 S 1</b>  |     |               |           |            |
|           | ID system with a 24 ID-card capacity  |     |               |           |            |
|           | ID cards incubation at 37 (+/-) degrees celcius   |     |               |           |            |
|           | Time and temperature are displayed (LCD) in English   |     |               |           |            |
|           | voltage: 110-240V, Frequency: 50-60HZ   |     |               |           |            |
|           | Weight: 6.0-6.5kg   |     |               |           |            |
|           | <b>LISS/Coombs</b>  |     |               |           |            |
|           | cards contain Polyspecific anti-human globulin (AHG) and anti-C3 to be used for antibody screening, crossmatch and direct antiglobulin test (DAT) |     |               |           |            |
|           | Complete Crossmatching  |     |               |           |            |
|           | A,B,DVIXAHG 48 profiles   |     |               |           |            |
|           | Autocontrol   |     |               |           |            |
| 2         | <b>Digital Centrifuge</b>   | 1   | PC            |           |            |
|           | <b>Specifications:</b>  |     |               |           |            |
|           | Motor: Brushless DC   |     |               |           |            |
|           | Timer (Min): 0-30 min. digital timer  |     |               |           |            |
|           | Display: LED digital display for time, speed, and RFC value   |     |               |           |            |
|           | Weight (kg) (N.W./G.W.) 1.0/21.0  |     |               |           |            |
|           | Overall Dimension (mm) (WxDxH) 390x505x209  |     |               |           |            |
|           | Speed (rpm): 200-4,000, 200-6,000   |     |               |           |            |
|           | Type 45" angle, swing out   |     |               |           |            |
|           | Max. RCF (xg) 2,325, 2,307, 5,151, 2,683  |     |               |           |            |
|           | Capacity 12x (5-15)ml, 24 x (5-15)ml, 6x50ml, 12x (5-15)ml, 16x(5-7)ml, 12x (10-15)ml, 4 x 50ml   |     |               |           |            |

|  |   |  |  |                         |                     |
|--|---|--|--|-------------------------|---------------------|
|  | Standard Accessory (5-7)ml and (10-15)ml tube adapters, 50ml tube adapter, (5-7)ml and 50ml tube adapters |  |  |                         |                     |
|  | <b>Features:</b>  |  |  |                         |                     |
|  | Metal housing is sturdy and durable.  |  |  |                         |                     |
|  | Maintenance-free brushless DC motor   |  |  |                         |                     |
|  | Auto-blancing system can extend motor life and reduce noise   |  |  |                         |                     |
|  | With safety interlock, it can prevent lid from being opened during operation                              |  |  |                         |                     |
|  | Forced fan air convection prevents the motor from overheating   |  |  |                         |                     |
|  | Precision digital timer can be set 30 minutes with 1sec increment.  |  |  |                         |                     |
|  | Digital speed setting is with 100rpm increment  |  |  |                         |                     |
|  | DSC-303SD can offer short runs by pushing the function key  |  |  |                         |                     |
|  | LED digital display enables users to monitor the time, speed and RFC values                               |  |  |                         |                     |
|  | Automatic brake system can offer rapid deceleration   |  |  |                         |                     |
|  | DSC-303SD features a see-through window   |  |  |                         |                     |
|  | The lid will be opened automatically once the rotor stops spinning  |  |  |                         |                     |
|  | Quality certificate ISO 9001, ISO 13485   |  |  |                         |                     |
|  | <b>***Nothing Follows***</b>  |  |  | <b>Grand Total(PhP)</b> | <b>1,285,000.00</b> |